




Accuracy

- GATF 2007 Spectro evaluation yielded the following:

IPA (Lab-Ref)



GATF 2007 Evaluation (1 = more than Delta E of 1.5)

													Patches with more than Delta 1.5
	White	Black	Cyan	Magneta	Yellow	Gray	Red	Green	Blue	Brown	Purple	Pastel	
SpectroEye	0	0	0	0	1	0	0	1	0	0	1	0	3
i1	0	0	0	0	0	0	0	0	0	0	1	0	1
X-Rite 530	0	0	0	0	0	1	0	0	1	0	1	0	3
X-rite939	0	0	1	0	0	0	0	0	1	0	1	0	3
DTP22	0	0	1	0	1	0	0	0	0	0	1	0	3
Techkon	0	0	0	0	0	0	0	0	0	0	0	0	0
Ihara S900	1	0	1	1	1	1	1	1	1	0	1	1	10